

The reason for cleaning an optical element is to improve performance. Proper materials, techniques and handling procedures should be used to minimize the risk of damage.

by Emily Kubacki, CVI Laser LLC

nyone who has worn glasses or driven a car knows the detrimental effects of dirt on the performance of a lens or window. The effects can be more serious in optics setups that involve laser radiation.

Contaminants, whether on a coating or on the substrate surface, can absorb laser energy, leading to damage to the coating or its interface with the substrate. Studies have shown that the cleaning process alone can cause variations in the laser damage threshold of a coated part by 25 to 75 percent. It is easy to see why cleaning, coating and packaging in a cleanroom environment are so important to both the lifetime and performance of optics.

Unfortunately, the photonics industry does not have established standards either for cleaning methods or even for the definition of what a "clean" optical element is. The only widely accepted principles are that all cleaning methods incur some handling risks and that excessive or unnecessary cleaning should be avoided to prevent limiting the lifetime of an optical component — in particular, coated optics.

Historically, optical components such as mirrors and beamsplitters have been cleaned by hand, using lint-free optical wipes and reagentgrade acetone or another liquid solvent such as methanol, ethanol, 97 percent pure isopropyl alcohol, methyl ethyl ketone (MEK) or methylene chloride (MEC). Some inorganic acids such as trichloroethylene (TCE), hydrofluoric acid (HF) and hydrochloric acid (HCL) may be used on uncoated silicon wafers, and nitric acid may be used on germanium substrates. Acidic solutions, however, should never be used on coated or uncoated zinc sulfide (ZnS) or zinc selenide (ZnSe) components.

Acetone is very good at dissolving grease, but it dries very quickly and always should be handled with acetone-impenetrable gloves. In general, isopropyl alcohol is a safe and effective cleaner — except for cleaning aluminum coatings. Because alcohol reacts with aluminum, it should never be used on protected or bare aluminum-coated mirrors. Methanol

and most acidic solutions can be toxic or damaging to optics or coatings if misused, so care should be taken to follow the instructions provided by the manufacturer.

Before cleaning, it generally is recommended that one first remove dust and other large particles of debris using a bulb blower, a camel hair brush or dry-nitrogen-propelled compressed air. Other pressurized gas products are not recommended because the propellant could be released onto the optical surface in the form of liquid droplets.

Three hand-cleaning techniques typically are used, depending on the type of optic and the level of cleaning required: the drop-and-drag method, the wipe method and immersion. For each, it is important to handle the optics with care. Using latex gloves or powder-free finger cots and always holding the component by the outer edges will help to prevent additional scratches, chips or fingerprints.

The drop-and-drag method, the most commonly used, is intended for light cleaning. A drop of methanol or reagent-grade acetone is placed in the center of a clean lens tissue, precision quality optical wipe or microfiber cloth; it should not be applied directly to the lens surface. The wet portion of the wipe is placed on the optic and slowly dragged across the surface, exerting no pressure. The surfaces of both the wipe and the optic should be nearly dry at the completion of the drag. If done properly, the residual solvent will have evaporated from the optical surface without leaving spots or wipe marks. In many cases, it is beneficial to clean the edges of the optic first to prevent dirt or excess polishing compound from being drawn onto the polished surface.

hniques

inorganic

ethylene

F) and hy-

y be used

s, and ni-

rmanium

ns, how-

on coated

iS) or zinc

ickly and

with ace-

n general,

and effec-

cleaning


use alco-

it should

d or bare

Methanol ARCH 2006

ts. dissolving The wipe method is recommended to remove fingerprints and other embedded residues on uncoated surfaces or durable coatings. This technique uses a clean lens tissue or wipe folded into a brush and grasped near the fold with a hemostat, fingers or tweezers. Several drops of solvent are placed on the fold, and the excess is shaken off. With gentle, uniform pressure, the brush is wiped slowly from one edge of the optic to the other



The wipe technique should involve smooth and continuous motions of the lens tissue across the optic.

and then discarded, even if a second wipe is required.

An alternative to this technique, which is particularly useful for infrared optics, uses sterile surgical-quality cotton balls or lint-free cotton swabs. Using only the pressure from the wipe or cotton ball, the optic can be wiped across in an "S" shape or along individual spokes, working from the center out to the edge. Each swab is discarded between wipes. It is important that the swab be dampened with acetone or alcohol before use and that the surface not

be scrubbed back and forth, which would increase the likelihood of surface scratches or damage to a coating.

The immersion technique can be used on most metallic coatings, such as protected silver, protected aluminum and protected gold. Each of these coatings, however, varies in hardness and durability and so should be treated with extreme caution. After the surface dust is removed, the optics can be immersed in acetone and rinsed in fresh solvent until clean. Bare metal coatings

Combinations to Avoid	
Alcohol	Aluminum coating or substrates
Acids (e.g., HF, HCL)	Coated optics Zinc sulfide (ZnS)
(e.g., 1 ii , 1 i C.)	Zinc selenide (ZnSe)
	VIS and UV optical glasses (e.g., BK 7, fused silica, CaF ₂ , MgF ₂)
	aan officialise all on a sea of the sea of t
MEC, MEK, TCE	Coated optics
All	Bare metal coatings
	(e.g., gold, copper, aluminum)
Ultrasonic cleaning	Metal coatings
	Soft optical glass (e.g., CaF,, MgF ₂ , SF11)





LightWise Smart Camera Series

• Line Scan cameras

• LW-ELIS-1024A-1394

- 10K lines per sec, 10MHz
- 7.8 x 125 Micron Pixels

o LW-SLIS-2048A-1394

- · 28K lines per sec, 60MHz
- 7 x 7 Micron Pixels

o LW-PE-2048-1394

- 19K lines per sec, 40MHz
- 14 x 14 Micron Pixels CCD

Area Cameras

o LW-WVGA-G-1394

- 72fps, Full Frame Global Shutter
- 752 x 480 Mono & Color
- 110db Dynamic Range!

o LW-1.3-S-1394 or LW-3-S-1394

- 30fps@1.3MP, 12fps@3MP
- Synchronous Shutter
- 1280 x 1024 or 2048 x 1536
- Color (-C) or Mono (-M)

o LW-1.3-G-1394-M

- 30fps, Full Frame Global Shutter
- 1280 x 1024 Mono only

o LW-6.6-S-1394 or LW-5-S-1394

- 6.6Mp 5fps@6.6MP, 89fps@VGA
- 6.6Mp 3002 x 2210 (-C) or (-M)
- 5Mp 12fps@5MP,30fps@VGA
- 5Mp 2592 x 1944 (-C) Color only
- Synchronous Shutter

Programmable Color Image Path w/ Digital Pan & Zoom + JPEG options, External Trigger & Strobe I/O, On-Board Image Buffer & FPGA's

For Free Info, Circle 59

Custom Camera Design

For more detailed specifications visit or call:

http://www.isgchips.com/photonics

If you register at this page and mention this advertisement ISG will send you a free gift.

sales@ISGChips.com 585-388-5220

Optics Cleaning


and some very soft coatings used for filters or deep-ultraviolet optics should not be cleaned or touched in any way.

Ultrasonic cleaning is a multistep aqueous process that typically is used on uncoated substrates prior to coating. Whereas hand cleaning primarily is done on coated optics and soft substrate materials — including calcium fluoride (CaF_2), magnesium fluoride (MgF_2) and flint

glasses such as SF11 — a combination of a degreaser, a mechanical scrub and/or an ultrasonic aqueous cleaning cycle generally is considered sufficient for "hard" types of glass; i.e., those with a Knoop coefficient ≥500, such as BK 7 and fused silica or quartz.

The process consists of a five- to seven-stage progression of rinses, sprays and washes in neutral detergents and high-purity deionized

Methods and Applications		
Brush or air blower	Large, loose dust particles on most coated or uncoated optics	
Drop-and-drag method	Light cleaning of dust and lint on most coated or uncoated optics	
Wipe method	Moderate cleaning of fingerprints or stains on substrates or "hard" coatings	
Immersion cleaning	Nonabrasive cleaning of "soft" and protected metal coatings	
Ultrasonic cleaning	Multistep cleaning cycle on "hard" glass types (prior to coating)	
CO ₂ cleaning	Spray cleaning on uncoated optics fibers, crystals and polymers	



Precision optics should be handled properly in cleaning. Here, the prism is held on the frosted surfaces.


water, followed by a hot air-dry and inspection in a class 100 or better cleanroom. It sometimes is preceded by a mechanical scrub or a vapor acetone or bromide degreaser step to optimize the efficiency of the clean-

ing cycle. Depending on the size and quantity of the optics, they may be held in stainless steel baskets or mounted in the appropriate coating fixtures to prevent recontamination from excess handling prior to coating.

Liquid CO₂ is a new technique that is used to remove oils and microscopic particles from optical waveguides, electro-optical devices, silicon wafers and a variety of biomedical, aerospace and semiconductor components. This process delivers a precisely controlled and purified spray alternated with warm air cy-

cles to the optical surface.

Because CO_2 is noncorrosive and relatively nontoxic, it is safer to use than many traditional solvents, but it requires nontraditional procedures and a controlled, moisture-free work environment and so may incur additional initial expenses. In the long term, however, it may turn out to be a less expensive and more effective means of achieving ultraclean surfaces, possibly resulting in coatings with higher damage thresholds. \Box



Ultrasonic cleaning is an automated multistep aqueous process for the cleaning of substrates prior to coating.

Meet the author

Emily Kubacki is the optics product manager at CVI Laser LLC in Albuquerque, N.M.; e-mail: emilyk@cvilaser.com.

High brightness LEDs and the tools to measure them.

Our experienced engineering staff will find the right LED for your application — from thru-hole to surface mount, visible to infrared — whether you're in scientific research, medical devices, commercial and architectural lighting or contract manufacturing for instrument makers.

We also have the tools to measure all of your LEDs. Our RPS900 Wideband Spectroradiometer includes the SpectriLight Tools™ analysis package for easy use, high accuracy, and multi-application measurement for a wide range of LEDs. It's ideal for production, inspection, development, and research applications.

For in stock LEDs and accurate LED measurement instruments, call us.



10 Technology Drive, Peabody, MA 01960 978-818-6180 / 978-818-6181 fax intl-light.com / gilway.com

Gilway

international Light

